

Corning Incorporated  
Life Sciences

Registered  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 1395-150

**Product Description:** PYREX® Bottle, Media, Storage, GL-45 Thread, 150 mL,  
with Orange PP Screw Cap and PP Drip Ring

**Lot Number Designation:**

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Component Materials:**

Glass - Type I Class A - Low-Expansion Borosilicate Glass,  
per ASTM E-438 (Standard Specification for Glasses in Laboratory Apparatus)  
Cap - Polypropylene (PP)  
Ring - Polypropylene (PP)

**Attributes:**

Capacity - 100 mL Approximate

**TSE/BSE:**

This product complies with the latest revision of EMEA/410/01, *Note for Guidance on Minimizing the Risk of Transmitting Animal Spongiform Encephalopathy Agents via Human and Veterinary Medicinal Product* by virtue that any component of the product which may include material of bovine origin has been processed per the specific conditions of section 6.4 of EMEA/410/01.

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

**Test/Procedure:**

Visual Attributes  
Dimensional  
Packaging

**Results:**

Pass  
Pass  
Pass

**Revision Date:**

17 JUN 2013



Arthur D. Birdsall  
Quality Assurance Manager

Rev No: 1