

Certificate of Compliance

Corning Incorporated - Life Sciences 271 County Route 64 Big Flats NY 14814 USA www.corning.com/lifesciences Refer to website for regional contact information.

Page: 1 / 1

Product Name : MEDIA STORAGE BOTTLE, 3500ML, SCREW CAP,1/CS

Catalog Number : 1395-3X Manufacture Date : 2019-12-11

Lot ID : 34419001

Expiration Date : Unconfirmed Validity

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

BSE/TSE - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass Packaging Inspection - Pass Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

ANDREW J (ANDY) FRANKS CMO Quality Manager

Creation Date: 2020-02-03 17:44:48