

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 2016
Lot Number: 06614003
Product Description: Corning® cap inserts for cryogenic vials, blue

Component Materials:
Cap - Virgin Polypropylene Heavy metal free (meets *CONEG* req.) color concentrate.

Pyrogens:
The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: *Bacterial endotoxins - Test methodologies, routine testing, and alternative to batch testing*. Results: less than 0.1 EU/mL

RNase/DNase Testing:
This product has been tested and is free of any detectable RNase/DNase contamination.

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:
This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Packaging	Pass

Lot Number Designation:
8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.



Viviana Niño
Quality Supervisor

Rev No: 3