

Certificate of Analysis

Description: HBSS, 1X
(Hank's Balanced Salt Solution)
with calcium & magnesium

Catalog #: 21-020-CV
Lot #: 24615006
Expiration Date: 09/18
Country of Manufacture: USA

Storage: 2-30°C

Notes:

Mediatech, Inc. products are prepared by a validated aseptic sterile filtration process to ensure that all products comply with the industry's Sterility Assurance Level (SAL) of at least 10^{-3} (i.e. demonstrating that a product produced in this manner has no more than 1 random contaminant per 1000 units).

This product was tested in accordance with currently-approved Mediatech, Inc. specifications and procedures. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

For IVD Use. Not for Human Therapeutic Use. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.


TEST PARAMETERS	SPECIFICATION	RESULT
pH	7.1 – 7.4	7.2
Osmolality	275 ± 15 mOsm/Kg H ₂ O	282 mOsm/Kg H ₂ O
Endotoxin	≤ 0.25 EU/mL @ FW	< 0.0010 EU/mL @ FW
Mycoplasma	Tested Negative	Negative
Sterility	Pass	Pass

For additional information on this Mediatech product, please contact our Technical Services department at (800) 235-5476 or via email at MITechServ@corning.com.

Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date: Augusta James 13 Oct 2015

Reviewed By/Date: Jacampo Ferrero / YF
14 Oct 2015



Mediatech, Inc.
A Corning Subsidiary
9345 Discovery Boulevard
Manassas, VA 20109
Main (703) 471-5955
Customer Service (800) 235-5476
www.cellgro.com