

## Certificate of Analysis

<b>Description:</b>	<b>Lymphocyte Separation Medium</b> Density = 1.077-1.080 g/mL @ 20°C	<b>Catalog #:</b>	25-072-CI
		<b>Lot #:</b>	09020013
		<b>Expiration Date:</b>	2021-09
<b>Storage:</b>	2-30°C	<b>Country of Manufacture:</b>	USA

**Notes:**

Mediatech, Inc. products are prepared by a validated aseptic sterile filtration process to ensure that all products comply with the industry's Sterility Assurance Level (SAL) of at least  $10^{-3}$  (i.e. demonstrating that a product produced in this manner has no more than 1 random contaminant per 1000 units).

This product was tested in accordance with currently approved Mediatech, Inc. specifications and procedures. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

For Research or Further Manufacturing Use Only Not for use in Diagnostic Procedures. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

TEST PARAMETERS	SPECIFICATION	RESULT
pH	6.0 – 9.0	6.5
Osmolality	290 ± 20 mOsm/kg H <sub>2</sub> O	288 mOsm/kg H <sub>2</sub> O
Density	1.077 – 1.080 g/mL @ 20°C	1.079 g/mL @ 20°C
Endotoxin	≤ 0.25 EU/mL @ 1:10	0.0026 EU/mL @ 1:10
Sterility	Pass	Pass
Separation of Lymphocytes	≥ 90%	99%

For further information, contact Corning Scientific Support at 800-492-1110 or [ScientificSupport@corning.com](mailto:ScientificSupport@corning.com).

Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date: Charity Stewart 29 Jun 2020 Reviewed By/Date: Racampo Lemos / YE 29 Jun 2020



Mediatech, Inc.  
A Corning Subsidiary  
9345 Discovery Boulevard  
Manassas, VA 20109  
1-800-492-1110  
[www.corning.com](http://www.corning.com)