Corning Incorporated Life Sciences

Registered ISO 9001:2008

Certificate of Compliance

Catalog Number: 2580

Lot Number: 14114012

Product Description: Costar ® Individual 1 x 8, high binding EIA/RIA strip, flat bottom, without lid

Component Materials:

Strips - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures.

Binding Characteristics:

This product has an approximate binding capacity of 500 ng IgG/cm^2 . The product has been tested for the attribute of well to well protein binding utilizing a modified ELISA to measure precision. Coefficient of variation must be $\leq 8\%$.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:Results:Visual AttributesPassPackagingPassFunctionalPass

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

Judith A. Gibbs

Quality Assurance Manager

fudith A. Glibbs

Rev No: 6