

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

# Certificate of Compliance

**Catalog Number:** 2580  
**Lot Number:** 14914023  
**Product Description:** Costar® Individual 1 x 8, high binding EIA/RIA strip, flat bottom, without lid  
**Component Materials:**  
Strips - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

**Binding Characteristics:**

This product has an approximate binding capacity of 500 ng IgG/cm<sup>2</sup>. The product has been tested for the attribute of well to well protein binding utilizing a modified ELISA to measure precision. Coefficient of variation must be ≤ 8%.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:


<b>Test/Procedure:</b>	<b>Results:</b>
Visual Attributes	Pass
Packaging	Pass
Functional	Pass

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Expiration Statement:**

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.



Judith A. Gibbs  
Quality Assurance Manager

Rev No: 6