| | Corning Incorporated Life Sciences | | | |
|----------------------------------|---------------------------------------|---|---------------------------|--|
| | Registered ISO 9001:2008 | | Certificate of Compliance | |
| (| Catalog Number: | 2580 | | |
| Lot Number: | | 17513034 | | |
| Product Description: | | Costar ® Individual 1 x 8, high binding EIA/RIA strip, flat bottom, without lid | | |
| Component Materials: Strips - | | Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and closures. | | |
| | | | | |

Binding Characteristics:

This product has an approximate binding capacity of 500 ng lgG/cm². The product has been tested for the attribute of well to well protein binding utilizing a modified ELISA to measure precision. Coefficient of variation must be $\leq 8\%$.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

| lest/Procedure: | Results: |
|-------------------|----------|
| Visual Attributes | Pass |
| Packaging | Pass |
| Functional | Pass |

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

Judith A. Gibbb

Judith A. Gibbs Quality Assurance Manager

Rev No: 6