

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance


Catalog Number: 3367
Lot Number: 01010001
Product Description: Costar® 96-well Serocluster®, round bottom, without lid

Component Materials:
Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

Sterilization:
This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.
Sterility Assurance Level: SAL 10⁻³

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Packaging	Pass



Judith A. Gibbs
Quality Assurance Manager

Rev No: 4