

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

# Certificate of Compliance

**Catalog Number:** 3367  
**Lot Number:** 16914042  
**Product Description:** Costar® 96-well Serocluster®, round bottom, without lid

**Component Materials:**  
Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.


**Sterilization:**

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*

Sterility Assurance Level: SAL 10<sup>-3</sup>

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

<b>Test/Procedure:</b>	<b>Results:</b>
Visual Attributes	Pass
Packaging	Pass



Judith A. Gibbs  
Quality Assurance Manager

Rev No: 4