

**Certificate of Analysis**

<b>Description</b>	<b>Fetal Bovine Serum</b>	<b>Catalog #:</b>	35-011-CV
	USDA Approved Source	<b>Lot #:</b>	35011115
	Heat Inactivated	<b>Expiration Date:</b>	05/2020
<b>Storage</b>	-25 to -15°C	<b>Country of Raw Material Origin:</b>	USA
		<b>Country of Manufacture:</b>	USA

**Notes:**

This product was tested in accordance with currently approved Mediatech, Inc. specifications and procedures. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

For Research Use Only. Not for Human Therapeutic Use. Not for Human or Animal Consumption. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

This product manufactured from material collected in US origin approved abattoirs. The serum was aseptically processed and filtered in a facility registered with medical devices and following guidelines of FDA.

TEST PARAMETERS	SPECIFICATION	RESULT
pH	7.8 ± 0.8	7.7
Osmolality	315 ± 35 mOsm/kg H <sub>2</sub> O	294 mOsm/kg H <sub>2</sub> O
Endotoxin	≤ 20 EU/mL	< 0.7 EU/mL
Mycoplasma	Tested Negative	Negative
Total Protein	3-5 g/dL	4 g/dL
Hemoglobin	< 30 mg/dL	14 mg/dL
Sterility	Sterile	Sterile
Performance Testing	Pass	Pass
Filtration	Triple 0.1 µm Filtered	Triple 0.1 µm filtered

For additional information on this Mediatech product, please contact our Technical Services department at (800) 235-5476 or via email at MTTechServ@corning.com.

Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date: Zara Brown 07 Jul 2015

Reviewed By/Date: Jacques Lemero / 45  
08 Jul 2015



Mediatech, Inc.  
A Corning Subsidiary  
9345 Discovery Boulevard  
Manassas, VA 20109  
Main (703) 471-5955  
Customer Service (800) 235-5476  
www.cellgro.com