CORNING

Certificate of Analysis

Heat Inactivated **Product:** Fetal Bovine Serum

Description: Regular,

South American Origin

Catalog Number: 35-089-CF, 35-089-CV Lot Number: 33820002

Date of Heat **Heat Inactivation** Individual Bottle 23 MAR 2021

Inactivation: Inactivation Type:

Product Prior to Heat Heat Inactivation 56°C (±2°C) for 30 **Testing Performed on:** Inactivation Temperature: minutes

Parent Catalog Parent Lot Number: 35-079 33820001 Number:

Date of Manufacture: 03 DEC 2020 **Date of Expiration: DEC 2025**

Country of Raw **Country of Final** Brazil

United Kingdom Material Origin: Processing:

Sterility Assurance Triple 0.1 µm Filtered -40°C to -10°C Storage: Level:

Specification Test Method Results Units Endotoxin USP <85> & EP 2.6.14 ≤ 20 < 0.4 EU/mL Hemoglobin ≤ 30 Spectrophotometric 7.54 mg/dL Mycoplasma Not Detected N/A Not Detected Osmolality Freezing point depression mOsm/kg H₂O 260 - 350305 рН 6.5 - 8.57.30 N/A pH Meter Sterility (Bacteria & Fungi) USP <71> & EP 2.6.1 No Growth No Growth N/A **Total Protein** 3.0 - 4.5USP <1057> & EP 2.5.33 3.8 g/dL Gamma-Glutamyl Transferase (GGT) ≤10 4.8 IU/L Cell Growth Performance Testing Passed Passed N/A Biochemical Testing - Proteins & Organic Compounds Alkaline Phosphatase IU/L As Reported 316 Cholesterol As Reported mg/dL 30.9 Creatinine mg/dL As Reported 3.5 Glucose As Reported 54.0 mg/dL Glutamic Oxaloacetic Transaminase (AST) As Reported IU/L 24.9 Immunoglobulin G (IgG) **ELISA** As Reported µg/mL 146.0 **Total Bilirubin** mg/dL As Reported 0.0 Triglyceride As Reported mg/dL 71.7 As Reported mg/dL Urea 37.8 Biochemical Testing - Electrophoretic Profile Normal Normal N/A Albumin As Reported g/dL 1.70 Alpha As Reported g/dL 1.53 Beta As Reported g/dL 0.41 Gamma As Reported g/dL 0.18 Biochemical Testing - Trace Metals Calcium As Reported 14.4 mg/dL

Magnesium	As Reported	3.4	mg/dL
Phosphorus	As Reported	9.3	mg/dL
Sodium	As Reported	138	mmol/L

Notes:

For additional information on this Mediatech product, please contact our Technical Services department at (800) 492-1110 or via email at scientificsupport@corning.com.

This product was tested in accordance with Mediatech, Inc. specifications and procedures current as of date of manufacture. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

Cell lines used in Cell Growth Performance Testing: MRC5, L929, C2C12, LLC-PK1, Vero, MDCK and Sp20-Ag14.

For research or further manufacturing uses only. Not for use as an excipient. Not for therapeutic or diagnostic uses. Not for human or animal consumption. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

Origin and BSE/TSE Statements:

The aforementioned lot was manufactured from fetal bovine blood collected from government approved harvest facilities located in Brazil, a country recognized by the United States Department of Agriculture - Animal and Plant Health Inspection Service (USDA-APHIS) and the World Organization for Animal Health (OIE) as having a negligible risk of Bovine Spongiform Encephalopathy (BSE), a bovine-specific Transmissible Spongiform Encephalopathy (TSE). All fetal bovine serum is derived from fetuses collected from healthy bovine cows and heifers that have passed ante and post mortem inspection and were found free of contagious diseases. This material was exported from Brazil accompanied by a Veterinary Health certificate. Once inside the European Union, the material was inspected at the Boarder Inspection Post and released by the issuance of a Community Veterinary Entry Document ensuring that the collection facility and the importer are both cleared as suitable exporters and importers within the EU Traceability system.

The following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date:

30 APR 2021

Reviewed By/Date:

Jacah For

Lot Number: 33820002