

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 356666
Lot Number: 12814035
Product Description: Corning® Biocoat™ 384 well plate, clear, collagen I coated, with lid

Component Materials:
Collagen I - Rat Tail Tendon
Resin - Clear Polystyrene
Lid - Clear Polystyrene

Bioburden:
This lot has been tested and found negative for Bacteria and Fungi.

Cell Attachment and Growth Characteristics:
This lot has been tested for the attribute of Cell Attachment and Growth utilizing human HT1080 cells, an attachment-dependent mammalian cell line in a serum free media.

Collagen I:
The plates are coated with Collagen I which is a rat tail collagen, coated via a proprietary method.

Storage Conditions:
For optimal performance, the unopened product should be stored in a dry place at 4-30 degrees C (Do not freeze).


Handling:
The product is ready to use and does not require washing of the coated surface prior to use.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

| | |
|------------------------|-----------------|
| Test/Procedure: | Results: |
| Visual Attributes | Pass |
| Bacteria & Fungi Free | Pass |
| Packaging | Pass |

Lot Number Designation:
8 Digit Lot Number: First 3 digits - Julian Date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Expiration Statement:
This product has an expiration date of 6 years from the date of manufacture for the attributes contained within this certificate.



Judith A. Gibbs
Quality Assurance Manager

Rev No: 1