# Corning Incorporated Life Sciences

Registered ISO 9001:2008

# Certificate of Compliance

Catalog Number: 3571BC

Product Description: Corning ® 384-well, tissue culture treated, black opaque plate, with lid, with barcode

labels

21614055

**Component Materials:** 

Lot Number:

Plate - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures. Black concentrate

Lid - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures.

Barcode Label - White Polypropylene base, acrylic adhesivve.

**Special Feature:** 

Barcode label (format 128) on A1 short & A1 long side and P24 short & P24 long side.

#### Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*Sterility Assurance Level: SAL 10<sup>-3</sup>

Pyrogens 8 1

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing. Results: ≤ 0.1 EU/mL ( ≤ 4EU/device)

## **Optical Characteristics:**

The product is made of opaque black polystyrene walls to minimize well to well crosstalk and background fluorescence and /or luminescence.

### Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of Cell Attachment and Growth utilizing an attachment-dependent mammalian cell line in a serum supplemented media.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:Results:Visual AttributesPassPackagingPassCell Culture TreatmentPass

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian Date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Expiration Statement:** 

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

Judith A. Gibbs

**Quality Assurance Manager** 

fuduth A. Glibbs

Rev No: 1