

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 3571BC

Product Description: Corning® 384-well, tissue culture treated, black opaque plate, with lid, with barcode labels

Component Materials:

- Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures. Black concentrate
- Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Barcode Label - White Polypropylene base, acrylic adhesive.

Special Feature:

Barcode label (format 128) on A1 short & A1 long side and P24 short & P24 long side.

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*

Sterility Assurance Level: SAL 10⁻³

Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing.* Results: ≤ 0.1 EU/mL (≤ 4EU/device)

Optical Characteristics:

The product is made of opaque black polystyrene walls to minimize well to well crosstalk and background fluorescence and /or luminescence.

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of Cell Attachment and Growth utilizing an attachment-dependent mammalian cell line in a serum supplemented media.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:


Test/Procedure:	Results:
Visual Attributes	Pass
Packaging	Pass
Cell Culture Treatment	Pass

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian Date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.



Judith A. Gibbs
Quality Assurance Manager

Rev No: 1