

Corning Incorporated - Life Sciences
1 Becton Circle
Durham NC 27712 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: Pipet 10mL Serol 50/pk 200/cs	Manufacture Date	: 2013-09-12
Catalog Number	: 357551		
Batch Number	: JW004		
Expiration Date	: 2018-09-11		

Quality System Compliance - BD Falcon™ products are manufactured under the ISO 9001: 2008 Standard, and the current FDA Quality System Regulation in Chapter 1, Title 21, Section 820 of the Code of USA Federal Regulations.

Products Labeled Non-Pyrogenic - Products labeled non-pyrogenic have been validated per FDA guidelines on LAL (Limulus Amebocyte Lysate) testing for medical devices and Company guidelines. The acceptance level for product is less than 0.1 EU/ml or 5 EU/device.

Sterilization - Product labeled as sterile is irradiated and dosimetrically released upon U.S. Association for the Advancement of Medical Instrumentation (AAMI) recommended practices in effect at the time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release. BD Biosciences - Discovery Labware products labeled sterile meet a minimum requirement of 10^{-3} SAL (Sterility Assurance Level).



Tara Rogers
Treyburn Quality Manager