

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 3591

Lot Number: 21312019

Product Description: Costar ® 96-well, medium binding EIA/RIA plate, flat bottom, without lid

Component Materials:

Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

Binding Characteristics:

This product has an approximate binding capacity of 250 ng IgG/cm². The product has been tested for the attribute of well to well protein binding utilizing a modified ELISA to measure precision. Coefficient of variation must be ≤ 8%.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:


Test/Procedure:	Results:
Visual Attributes	Pass
Packaging	Pass

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.



Judith A. Gibbs
Quality Assurance Manager

Rev No: 5