

CORNING LIFE SCIENCE (WUJIANG) CO LTD - Life Sciences  
T03/17 BUILDING, 1801 PANG JIN ROAD  
WUJIANG, JIANGSU PROVINCE 215200 CHN  
www.corning.com/lifesciences  
Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b>	: PLATE,96WL,TCT,FB,PS,W/LID,S, CNR NTCH		
<b>Catalog Number</b>	: 3596	<b>Manufacture Date</b>	: 2023-08-21
<b>Lot ID</b>	: 23323601		
<b>Expiration Date</b>	: 2026-08-20		

**Quality Management System** - Complies with the current version of the EN ISO 13485 Standard.

**Animal Content** - Product does not contain materials of animal origin.

**Non-Pyrogenic** - Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing" and USP <85>, "Bacterial Endotoxins Test". The acceptance level for product is  $\leq 0.10$  EU/ml or  $\leq 4$  EU/device.

**USP Class VI Testing** - All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

**Sterility** - Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of  $10^{-3}$ .

**Tissue Culture** - Tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line. A minimum of 95% confluency is required for acceptance.

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Packaging Inspection - Pass
- Cell Attachment & Growth Treatment Verification - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

*Thomas Liu*

Huihao (Thomas) Liu  
Wujiang Quality Manager