

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

# Certificate of Compliance

**Catalog Number:** 3696  
**Lot Number:** 21312033  
**Product Description:** Costar® 96-well cell culture multiple well plate, half area plate, with low evaporation lid

**Component Materials:**  
Plate/Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

**Sterilization:**  
This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.  
Sterility Assurance Level: SAL 10<sup>-3</sup>

**Pyrogens:**  
The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing*.  
Results: ≤ 0.1 EU/mL ( ≤ 4EU/device)

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**  
This product does not contain materials of animal origin.


**Cell Attachment and Growth Characteristics:**  
The product has been tested for the attribute of Cell Attachment and Growth utilizing an attachment- dependent mammalian cell line in a serum supplemented media.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

<b>Test/Procedure:</b>	<b>Results:</b>
Visual Attributes	Pass
Cell Culture Treatment	Pass
Packaging	Pass

**Lot Number Designation:**  
8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Expiration Statement:**  
This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.



Judith A. Gibbs  
Quality Assurance Manager

Rev No: 15