

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 3696

Product Description: Costar® 96-well cell culture multiple well plate, half area plate, with low evaporation lid

Component Materials:
Plate/Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

Sterilization:
This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*
Sterility Assurance Level: SAL 10⁻³

Pyrogens:
The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing.* Results: ≤ 0.1 EU/mL (≤ 4EU/device)

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:
This product does not contain materials of animal origin.


Cell Attachment and Growth Characteristics:
The product has been tested for the attribute of Cell Attachment and Growth utilizing an attachment- dependent mammalian cell line in a serum supplemented media.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Cell Culture Treatment	Pass
Packaging	Pass

Lot Number Designation:
8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Expiration Statement:
This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.



Judith A. Gibbs
Quality Assurance Manager

Rev No: 15