

Corning Incorporated - Life Sciences  
6507 South 400 West  
Salt Lake City UT 84107 USA  
www.corning.com/lifesciences  
Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b>	: TIPS,10UL EXT LENGTH NAT GRAD LOW BIND IN HINGED RACKS,ST		
<b>Catalog Number</b>	: 4150	<b>Manufacture Date</b>	: 2018-12-27
<b>Lot ID</b>	: 36118000		
<b>Expiration Date</b>	: 2022-01-27		

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**BSE/TSE** - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

**Non-Pyrogenic** - Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing" and USP <85>, "Bacterial Endotoxins Test". The acceptance level for product is  $\leq 0.02$  EU/ml or  $\leq 0.8$  EU/device.

**USP Class VI Testing** - All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

**Human DNA Free** - Tested by PCR method and found to be free of detectable human DNA contamination. The assay detection limit is hDNA 5pg.

**DNase/RNase Free** - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is  $10^{-5}$  Kunitz units/ul for DNase and  $10^{-9}$  Kunitz units/ul for RNase.

**Sterility** - Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of  $10^{-3}$ .

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass  
Packaging Inspection - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Jeff Dunn  
Quality Manager