

Corning Incorporated  
Life Sciences

Registered  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 4152  
**Lot Number:** 32310000  
**Product Description:** DeckWorks™ Low Binding Pipet Tips, 1-300 µL, Graduated, Polypropylene, Natural, Sterile, in Hinged Rack, 96 Tips/Rack, 10 Racks/Pack (960 Tips), 4 Packs/Case (3840 Tips)

**Lot Number Designation:**

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date; Next 2 digits - last 2 digits of year of manufacture; Last 3 digits - Batch identification.

**Component Materials:**

- Tip - Polypropylene, virgin, color natural.
- Rack - Polypropylene, lid color natural, base color orange. Heavy metal free color concentrate (meets CONEG requirements).
- Deck - Polypropylene, color gray. Heavy metal free color concentrate (meets CONEG requirements).

**RNase/DNase Testing:**

This product has been tested to confirm that no detectable RNase/DNase contamination is present.

**Pyrogens:**

This product has been tested and has met the criteria established in *Guideline on Validation of the Limulus Amebocyte Lystate Test as an End-Product Endotoxin Test for Human and Parenteral Drugs, Biological Products and Medical Devices* (FDA, December, 1987). Results less than 0.1 EU/mL.

**PCR Certification:**

This product has been tested to confirm that no detectable PCR inhibitors are present.

**Sterilization:**

This product is sterilized using a gamma radiation method compliant with the requirements of ANSI/AAMI/ISO 11137-2:2006 to achieve a Sterility Assurance Level (SAL) of  $10^{-3}$  for the product.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Packaging	Pass



Arthur D. Birdsall  
Quality Assurance Manager

Rev No: 1