

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 430625

**Lot Number:** 23605002

**Product Description:** Corning® 150 mL bottle top filter, 33 mm neck size, cellulose acetate membrane, 0.45 µM pore size

**Expiration Date:**  
An expiration date exists for this product. Please refer to the case label and the filter funnel for the expiration date for this product.

- Component Materials:**
- Funnel & Funnel Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
  - Funnel Adapter & Hose Connector - Virgin High Density Polyethylene, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate.
  - Membrane - Cellulose Acetate, 0.45 µM pore size
  - Gasket - Polyethylene Foam liner
  - Filter Plug - Cellulose Acetate

**Sterilization:**  
This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*  
Sterility Assurance Level: SAL 10<sup>-5</sup>

**Pyrogens:**  
The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987).*  
Results: less than 0.1 EU/mL

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

| <b>Test/Procedure:</b> | <b>Results:</b> |
|------------------------|-----------------|
| Leak Test              | Pass            |
| Visual Attributes      | Pass            |
| Packaging              | Pass            |



Wendy Loitsch  
Quality Manager

Rev No: 6