

Certificate of Compliance

Corning Incorporated - Life Sciences Avenida Industrial del Norte S/N Parque Industrial del Norte Reynosa TAMP 88736 MEX www.corning.com/lifesciences Refer to website for regional contact information.

Page: 1 / 1

Product Name	:	CRYO BOX 81,1-2ML,NS,BK,	5/10	
Catalog Number	:	431119	Manufacture Date	: 2023-02-16
Lot ID	:	04723002		
Expiration Date	:	Unconfirmed Validity		

Quality Management System - Complies with the current version of the EN ISO 13485 Standard and the FDA CFR 21 Part 820, current Good Manufacturing Practices (cGMP).

BSE/TSE - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

USP Class VI Testing - All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass Functional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

Baudel Ruiz Quality Manager