	Corning Incorporated Life Sciences Registered ISO 9001:2008		Certificate of Compliance	
Ca	talog Number:	431406		
Lot Number:		33406001		
Product Description:		Corning® 250 mL disposable Baffled Erlenmeyer flask, plug seal cap		
closures Cap - Virgin P		- Virgin Polypre	rbonate, meets <i>USP, Class VI</i> requirements for plastic containers and opylene, meets <i>USP, Class VI</i> requirements for plastic containers and avy metal free (meets <i>CONEG</i> req.) color concentrate.	
04.				

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.* Sterility Assurance Level: SAL 10⁻⁶

Pyrogens:

The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Parenteral Drugs,Biological Products and Medical Devices (FDA, December, 1987).* Results: less than 0.1 EU/mL

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics: **Test/Procedure: Results:**

Test/Procedure:	Resul
Leak Test	Pass
Visual Attributes	Pass
Packaging	Pass

Wendy Lotsch

Wendy Loitsch Quality Manager

Rev No: 2