

Certificate of Compliance

Corning Incorporated - Life Sciences 275 River Street Oneonta NY 13820 USA www.corning.com/lifesciences Refer to website for regional contact information

Page: 1 / 1

Product Name : SPIN-X UF, 6mL CONCENTRATOR,50K CUTOFF,NS,IND,1/25

Catalog Number : 431485 Manufacture Date : 2019-10-29

Lot ID : 30219006 Expiration Date : 2021-10-29

Quality Management System - Complies with the current version of the ISO 9001 Standard and the FDA CFR 21 Part 820, current Good Manufacturing Practices (cGMP).

Membrane Recovery Efficiency - Tested at 10,000 xg, using a molecular weight marker of appropriate cutoff for the level of concentrator. Acceptance is a minimum recovery of 20% filtrate.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass Packaging Inspection - Pass Integrity Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

DAVID P (DAVE) KIERU Quality Manager

Creation Date: 2022-11-29 13:36:57