

Certificate of compliance

Date of certificate: 2019 - 10 - 14

Page: 1/1

CORNING GOSSELIN SAS 123 Rue de Caëstre - CS40019 59529 HAZEBROUCK CEDEX BORRE - FRANCE

**Product description** 

Octagonal bottle 30ml PET na graduated H62 Ø18 & tamper-evident screw cap na, 24/filmed pack, STERILE R

Catalogue number 431729

Manufacturing date 2017/09/25

**Lot number** *662339* 

## Raw material statement

The Substances Bisphenol A, BSE / TSE, latex, and phthalates are not used or intentionally incorporated in the manufacture of the raw materials or final product.

## **Sterility Assurance Level**

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-6

Leakproof - Not Applicable

**Non-Pyrogenic -** Tested and met the criteria established in the current version of United States Pharmacopeia (USP) Chapter <85> "Bacterial Endotoxins Test." The acceptance level for product is less than 0.1 EU/ml.

**USP Class VI Testing -** All material resin is tested, qualified and shown to be non-toxic as established in the standards USP Class VI Chapter <87>, "Biological reactivity Tests, in Vitro" and Chapter <88>, "Biological Reactivity Tests, in Vivo."

**Expiration date** 2020/09/01

## **Expiration statement**

This product is designed to have and expiration date of 3 years from the date of manufacture regarding sterility.

## Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

Cédric MAES

Plant Quality Manager, Borre & Chorges

maesc@corning.com