

Certificate of compliance

Date of certificate: 2018 - 10 - 03

Page: 1/1

CORNING GOSSELIN SAS 123 Rue de Caëstre - CS40019 59529 HAZEBROUCK CEDEX BORRE - FRANCE

Product description

Octagonal bottle 60ml PET naOctagonal bottle 60ml PET na graduated H62 Ø18 & tamper-evident scre

Catalogue number

431730

Lot number

629206

Sanitary Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-6

Manufacturing date

2016/11/16

Expiration date

2019/11

Expiration statement

This product is designed to have and expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

Cédric MAES

Plant Quality Manager, Borre & Chorges

maesc@corning.com