

Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 431730

Product Description: Corning® Storage Bottle, 60 mL, Octagonal, PET, with 18 mm HDPE Screw Cap, Sterile, 24/Tray, 96/Case

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

Component Materials:

- Bottle - PET, clear, non-cytotoxic, non-hemolytic, meets USP, Class VI requirements for plastic containers and closures.
- Cap - HDPE, natural, non-cytotoxic, non-hemolytic, meets USP, Class VI requirements for plastic containers and closures.

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137.
Sterility Assurance Level (SAL): 10⁻⁶

Pyrogens:

This product has been evaluated per USP and meets the requirements for classification as non-pyrogenic.
Results less than 0.1 EU/mL.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:
Visual Attributes
Packaging

Results:
Pass
Pass

Revision Date:
20 NOV 2013



Arthur D. Birdsall
Quality Assurance Manager

Rev No: 1