

Corning Incorporated  
Life Sciences

Registered  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 431731

**Lot Number:** 15811000

**Product Description:** Corning® Storage Bottle, 125 mL, Octagonal, PET, with 31.7 mm HDPE Screw Cap, Sterile, 24/Tray, 48/Case

**Expiration Statement:**

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

**Component Materials:**

Bottle - PET, clear, non-cytotoxic, non-hemolytic, meets USP, Class VI requirements for plastic containers and closures.

Cap - HDPE, natural, non-cytotoxic, non-hemolytic, meets USP, Class VI requirements for plastic containers and closures.

**Sterilization:**

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137.

Sterility Assurance Level (SAL):  $10^{-6}$

**Pyrogens:**

This product has been evaluated per USP and meets the requirements for classification as non-pyrogenic.

Results less than 0.1 EU/mL.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

**Test/Procedure:**

Visual Attributes  
Packaging

**Results:**

Pass  
Pass



Arthur D. Birdsall  
Quality Assurance Manager

Rev No: 2