# CORNING

# Certificate of compliance

Date of certificate : 2022 - 03 - 15

Page 1

CORNING GOSSELIN SAS 123 Rue de Caëstre - CS40019 59529 HAZEBROUCK CEDEX BORRE - FRANCE www.corning.com/lifesciences Refer to website for regional contact information

## **Product description**

Octagonal bottle, 125 mL PET na graduated H115Ø31 & tamper-evident screw cap na, 24/filmed pack, STERILE R

Catalogue number 431731

Lot number 796766

# Manufacturing date 2021/09/09

Quality Management System : Complies with the current version of the ISO9001 standard

#### **Raw material statement**

The Substances Bisphenol A, BSE / TSE, latex, and phthalates are not used or intentionally incorporated in the manufacture of the raw materials or final product.

#### **Sterility Assurance Level**

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-6

**Non-Pyrogenic** - Tested and met the criteria established in the current version of United States Pharmacopeia (USP) Chapter <85> "Bacterial Endotoxins Test." The acceptance level for product is less than 0.1 EU/ml.

**USP Class VI Testing -** All material resin is tested, qualified and shown to be non-toxic as established in the standards USP Class VI Chapter <87>, "Biological reactivity Tests, in Vitro" and Chapter <88>, "Biological Reactivity Tests, in Vivo."

#### **Expiration date**

2024/09/01

## **Expiration statement**

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

## **Performance testing (AQL)**

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release

Hugues Vanderschelden Plant Quality Manager, Borre vandersch2@corning.com