Corning Incorporated Life Sciences			
Registered ISO 9001:2008		Certificate of Compliance	
Catalog Number:	4715		
Lot Number:	19613001		
Product Description:	Corning ® R	Corning ${f B}$ Redi-Tip General Purpose Pipette Tip, 100 - 1000 μL , 100 tip hinged rack	
Component Materials: Tip	Virgin Polypro	Virgin Polypropylene with blue colorant, meets USP, Class VI requirements for plastic containers and closures. Heavy metal free (meets CONEG req.) color concentrate.	
Rack -	Polypropylene with orange colorant, meets <i>USP</i> , <i>Class VI</i> requirements for plastic containers and closures. Heavy metal free (meets <i>CONEG</i> req.) color concentrate. Polypropylene, meets <i>USP</i> , <i>Class VI</i> requirements for plastic containers and closures.		
Lid -			
Special Feature:			

The entire unit is autoclavable.

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Pyrogens:

The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: Bacterial Endotoxins - Test methodologies, routine tseting, and alternative to batch testing. Results: less than 0.1 EU/mL

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics: Test/Procedure: Results:

Test/Procedure:	Resu
Visual Attributes	Pass
Packaging	Pass

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification

Viviana Niño Quality Supervisor

Rev No: 3