

Certificate of Complaince

836 North Street - Life Sciences Tewksbury MA 01876 USA www.corning.com/lifesciences Refer to website for regional contact information.

Page: 1 / 1

Product Name : Corning Pluronics Sacrificial Ink 2.7mL/ Syringe

Catalog Number : 6157 Manufacture Date : 2022-07-30

Lot ID : 9198

Expiration Date : 2024-07-30

Quality Management System - Complies with the current version of the ISO 9001 Standard.

 ${f Sterility}$ - Tested for the presence of bacteria and fungi according to EP 6.2.1 "Sterility".

Viscosity - Rheometry tested at room temperature and demonstrated to be 860±150 Pa·s.

pH - 6.5-7.4 pH.

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

ANDREW J (ANDY) FRANKS CMO Quality Manager

Creation Date: 2023-04-17 00:52:23