

Certificate of Compliance

836 North Street - Life Sciences Tewksbury MA 01876 USA www.corning.com/lifesciences Refer to website for regional contact information.

Page: 1 / 1

Product Name :Ascent FBR 5m2 Bioreactor Consumable

Catalog Number :6973 Manufacture Date : 2022-07-15

Lot ID :25499827 Expiration Date :2023-07-15

Quality System:

Complies with the current version of the ISO 9001 Standard.

BSE/TSE:

Product is not manufactured with materials of animal origin.

Non-Pyrogenic:

All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

Sterilization:

Product has been Gamma irradiated

Quality Control:

Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass

Packaging Inspection - Pass

Leak Test - Pass

ANDREW J (ANDY) FRANKS CMO Quality Manager

Creation Date: 2022-08-26 14:48:MAN