

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 9016
Lot Number: 25813004
Product Description: 2 mL non-printed pipette, individual, paper/plastic wrap
Component Materials:
Pipette - Polystyrene, meets *USP Cytotoxicity Test <87>*.

Sterilization:

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*.


Sterility Assurance Level: SAL 10⁻³

Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing*. Results: ≤ 0.1 EU/mL (≤ 4EU/device)

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Packaging	Pass



Judith A. Gibbs
Quality Assurance Manager

Rev No: 4