

Certificate of Compliance

Description: 10-20mL Cryo Bag Catalog #: 91-200-88
Lot #: 157165
Expiration Date: 06/2021
Storage 10 to 34°C Country of Manufacture: USA

Notes: This is to certify that this product has been manufactured in a manner compliant with current Good Manufacturing Practices (cGMPs), FDA regulation 21 CFR part 820 and ISO 13485 Quality Standard.

This product is intended for Research and Manufacturing Uses. Not for Human Therapeutic Use. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

Physical Inspection: Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.
Packaging: Inspection for packaging integrity, accurate labeling and correct product configuration.

Sterilization: The product has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 Sterilization of healthcare products - Requirements for validation and routine control - Radiation sterilization. Sterility Assurance Level: SAL 10^-6

For further information, contact Corning Scientific Support at 800-492-1110 or ScientificSupport@corning.com

Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date: [Signature] 12 Jul 2019

Reviewed By/Date: [Signature] 12 Jul 2019



Mediatech, Inc.
A Corning Subsidiary
9345 Discovery Boulevard
Manassas, VA 20109
1-800-492-1110
www.corning.com