

Certificate of Compliance

Description	22L Rocker Bag	Catalog #:	91-200-92
		Lot #:	95714-0000
		Expiration Date:	Aug 2019
Storage	10 to 34°C	Country of Manufacture:	USA

Notes:

This is to certify that this product has been manufactured in a manner compliant with current Good Manufacturing Practices (cGMPs), FDA regulation 21 CFR part 820 and ISO 13485 Quality Standard. This product has been inspected and tested in accordance with currently approved Mediatech, Inc. applicable material specification and drawing requirements.

This product is intended for *Further Manufacturing/Processing Use*. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

Physical Inspection: Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for packaging integrity, accurate labeling and correct product configuration.

Sterilization: The product has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products – Requirements for validation and routine control – Radiation sterilization*.
Sterility Assurance Level: SAL 10^{-6}

For additional information on this Mediatech product, please contact our Technical Services department at (800) 235-5476 or via email at MTTechServ@corning.com.

Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date: Zara Brown 14 Sep 2016

Reviewed By/Date: [Signature] 15 Sep 2016

CORNING

Mediatech, Inc.
A Corning Subsidiary
9345 Discovery Boulevard
Manassas, VA 20109
Main (703) 471-5955
Customer Service (800) 235-5476
www.cellgro.com