

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE

Date of certificate : 2021 - 08 - 10

Page : 1/1

### **Product description**

*Dish 60x15 0Vent Sterile DoubleOute rBag 15/B 1620/CS*

### **Catalogue number**

*BP50-02*

### **Manufacturing date**

*2021/04/13*

### **Lot number**

*782041*

### **Sterility Assurance Level**

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10<sup>-6</sup>

### **Expiration date**

*2024/04/01*

### **Expiration statement**

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

### **Performance testing (AQL)**

*Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.*



Hugues Vanderschelden  
Plant Quality Manager, Borre  
vandersch2@corning.com