

Certificate of compliance

Date of certificate: 2021 - 11 - 22

Page: 1/1

CORNING GOSSELIN SAS 123 Rue de Caëstre - CS40019 59529 HAZEBROUCK CEDEX BORRE - FRANCE

**Product description** 

Dish 60x15 0Vent Sterile DoubleOute rBag 15/B 1620/CS

Catalogue number

BP50-02

Lot number

787389

**Manufacturing date** 2021/06/01

## **Sterility Assurance Level**

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-6

## **Expiration date**

2024/06/01

## **Expiration statement**

This product is designed to have and expiration date of 3 years from the date of manufacture regarding sterility.

## Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

D

Hugues Vanderschelden Plant Quality Manager, Borre vandersch2@corning.com