

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information

Date of certificate : 2023 - 05 - 03

Page 1

### Product description

*Dish 60x15 6 Vents Aseptic DoubleOu terBag 15/B 1620/CS*

### Catalogue number

*BP53-06*

### Manufacturing date

*2023/04/22*

### Lot number

*853393*

### Quality Management System

Complies with the current version of the ISO9001 standard.

### Sterility Assurance Level

ASEPTIC products are manufactured in a controlled environment, without human intervention during the manufacturing process. Finished products with an Aseptic claim are not subjected to sterilization and do not make a claim to have a Sterility Assurance Level.

### Expiration statement

*N/A*

### Performance testing (AQL)

*Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.*

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Hugues Vanderschelden  
Plant Quality Manager, Borre  
[vandersch2@corning.com](mailto:vandersch2@corning.com)