

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE

Date of certificate : 2018 - 10 - 03

Page : 1/1

Product description

Bottle Square HDPE 500mL 37mmRedCapWad AssembledSterile100CS

Catalogue number

FC500-03

Lot number

641541

Sanitary Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10⁻³

Manufacturing date

2016/12/02

Expiration date

2019/12

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.



Cédric MAES
Plant Quality Manager, Borre & Chorges
maesc@corning.com