

Certificate of compliance

Date of certificate: 2019 - 10 - 14

Page: 1/1

CORNING GOSSELIN SAS 123 Rue de Caëstre - CS40019 59529 HAZEBROUCK CEDEX BORRE - FRANCE

Product description

Bottle Square HDPE 500mL 37mmRedCap Wad AssembledSterile100CS

Catalogue number *FC500-03*

Lot number 718829

Manufacturing date 2019/07/10

Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-3

Expiration date

2022/07/01

Expiration statement

This product is designed to have and expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

Cédric MAES

Plant Quality Manager, Borre & Chorges

maesc@corning.com