Certificate of Compliance



Corning Incorporated - Life Sciences Avenida Industrial del Norte S/N Parque Industrial del Norte Reynosa TAMP 88736 MEX www.corning.com/lifesciences Refer to website for regional contact information.

Page: 1 / 1

Product Name	:	PlateMax Ultra Clear	permanent heat	sealing	film for qPCR, 1
Catalog Number	:	HS-100-QPCR	Manufacture	Date :	2020-06-13
Lot ID	:	16320033			
Expiration Date	:	2023-06-13			

Quality Management System - Complies with the current version of the ISO 9001 Standard and the ISO 13485 Standard.

 $DNase/RNase\ Free$ - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is 10^{-5} Kunitz units/ul for DNase and 10^{-9} Kunitz units/ul for RNase.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass Functional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

JOSE DOMINGUEZ Quality Manager