

Axygen Incorporated - Life Sciences
3321 Central Avenue
Union City CA 94587 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: PlateMax Ultra Clear permanent heat sealing film for qPCR, 1		
Catalog Number	: HS-100-QPCR	Manufacture Date	: 2014-10-21
Lot ID	: 29414000		
Expiration Date	: 2016-10-20		

Quality Management System - Complies with the current version of the ISO 9001 Standard and the ISO 13485 Standard.

Product Claim -

Lot Claim -

DNase/RNase Free - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is 10^{-5} Kunitz units/ul for DNase and 10^{-9} Kunitz units/ul for RNase.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass
Functional Test - Pass



Kelly C McMullin
Quality Manager