

## **Certificate of Compliance**

Corning Incorporated - Life Sciences
Avenida Industrial del Norte S/N
Parque Industrial del Norte
Reynosa TAMP 88736 MEX
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name : PlateMax Ultra Clear permanent heat sealing film for qPCR, 1

Catalog Number : HS-100-QPCR Manufacture Date : 2017-11-27

Lot ID : 31017001 Expiration Date : 2020-11-26

Quality Management System - Complies with the current version of the ISO 9001 Standard and the ISO 13485 Standard.

<code>DNase/RNase Free</code> - Tested by nuclease assay method and found to be free of detectable <code>DNase/RNase</code> contamination. The assay detection limit is  $10^{-5}$  Kunitz units/ul for <code>DNase</code> and  $10^{-9}$  Kunitz units/ul for <code>RNase</code>.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass Functional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

JOSE DOMINGUEZ Quality Manager

Creation Date: 2017-12-13 22:39:41