

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE
www.corning.com/lifesciences
Refer to website for regional contact information

Date of certificate : 2022 - 05 - 30

Page 1

Product description

AutoclaveBag 3L PP 40µm 500/CS

Catalogue number

SA03-01

Manufacturing date

2022/03/10

Lot number

803542

Lots : DATE FAB 02/2022

Quality Management System : Complies with the current version of the ISO9001 standard

Sterility Assurance Level

Non sterile

Expiration statement

N/A

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Hugues Vanderschelden
Plant Quality Manager, Borre
vandersch2@corning.com