

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information

Date of certificate : 2022 - 12 - 05

Page 1

### Product description

*AutoclaveBag 9L PP 40µm 500/CS*

### Catalogue number

*SA09-01*

### Lot number

*808426*

*Lots : FAB : MAI 2022*

### Manufacturing date

*2022/06/10*

**Quality Management System** : Complies with the current version of the ISO9001 standard

### Sterility Assurance Level

Non sterile

### Expiration statement

*N/A*

### Performance testing (AQL)

*Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.*

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Hugues Vanderschelden  
Plant Quality Manager, Borre  
[vandersch2@corning.com](mailto:vandersch2@corning.com)