

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE
www.corning.com/lifesciences
Refer to website for regional contact information

Date of certificate : 2023 - 05 - 26

Page 1

Product description

AutoclaveBag 9L PP 40µm 500/CS

Catalogue number

SA09-01

Lot number

835798

Lots : DATE FAB 12/2022

Quality Management System

Complies with the current version of the ISO9001 standard.

Manufacturing date

2023/01/19

Sterility Assurance Level

Non sterile

Expiration statement

N/A

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Hugues Vanderschelden
Plant Quality Manager, Borre
vandersch2@corning.com