

Certificate of compliance

Date of certificate: 2023 - 06 - 14

Page 1

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE
www.corning.com/lifesciences
Refer to website for regional contact information

Product description

AutoclaveBag 33L PP 40µm 500/CS

Catalogue number

SA33-01

Lot number

839841

Lots : date fab 01/2023, FAB : 01/2023

Quality Management System

Complies with the current version of the ISO9001 standard.

Manufacturing date 2023/04/12

Sterility Assurance Level

Non sterile

Expiration statement

N/A

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release

Hugues Vanderschelden Plant Quality Manager, Borre vandersch2@corning.com